

# CardioChek® Plus Quality Control and Optics Log

## Multiple Analyzer (v 1.09 or above)

<b>Account Name/Location:</b>		<b>Multi-Chemistry Control</b>		Lot #:
<b>Event ID:</b>		<b>Result Range</b>		<b>Level 1</b>
<b>PTS Panels® Test Strips:</b>				<b>Level 2</b>
<b>PTS Panels® Lipid strips:</b>		Total Cholesterol		
<b>PTS Panels® eGLU strips:</b>		HDL		
Other:		Triglycerides		
		GLU/eGLU		

**Important:**

1. Use the gray CardioChek check strips to conduct an Optics Test prior to running controls.
2. Do not use PTS Panels® Control Solution beyond the expiration date on the vial. Open product is stable for 10 months at room temperature or until the expiration date, whichever comes first.

**Serial #:** \_\_\_\_\_ **Optics Check:**  Pass  Fail

		Level 1					Level 2					Repeat: <input type="checkbox"/> Level 1 <input type="checkbox"/> Level 2				
Date	Initial	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail
						<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F

Notes:

**Serial #:** \_\_\_\_\_ **Optics Check:**  Pass  Fail

		Level 1					Level 2					Repeat: <input type="checkbox"/> Level 1 <input type="checkbox"/> Level 2				
Date	Initial	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail
						<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F

Notes:

**Serial #:** \_\_\_\_\_ **Optics Check:**  Pass  Fail

		Level 1					Level 2					Repeat: <input type="checkbox"/> Level 1 <input type="checkbox"/> Level 2				
Date	Initial	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail
						<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F

Notes:

**Serial #:** \_\_\_\_\_ **Optics Check:**  Pass  Fail

		Level 1					Level 2					Repeat: <input type="checkbox"/> Level 1 <input type="checkbox"/> Level 2				
Date	Initial	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail
						<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F

Notes:

**Problem and Corrective Action Log:**

For any questions or concerns, call your event manager, follow the user guides and package inserts provided with the products or call PTS Diagnostics Customer Service: 1 (877) 870-5610 or +1 (317) 870-5610 for toll-free calls



# CardioChek® Plus Quality Control and Optics Log

## Multiple Analyzer (v 1.09 or above)

<b>Account Name/Location:</b>		<b>Multi-Chemistry Control</b>		Lot #:
<b>Event ID:</b>		<b>Result Range</b>		<b>Level 1</b>
<b>PTS Panels® Test Strips:</b>				<b>Level 2</b>
<b>PTS Panels® Lipid strips:</b>		Total Cholesterol		
<b>PTS Panels® eGLU strips:</b>		HDL		
Other:		Triglycerides		
	Lot #:	GLU/eGLU		

### Important:

1. Use the gray CardioChek check strips to conduct an Optics Test prior to running controls.
2. Do not use PTS Panels® Control Solution beyond the expiration date on the vial. Open product is stable for 10 months at room temperature or until the expiration date, whichever comes first.

**Serial #:** \_\_\_\_\_ **Optics Check:**  Pass  Fail

		Level 1					Level 2					Repeat: <input type="checkbox"/> Level 1 <input type="checkbox"/> Level 2				
Date	Initial	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail
						<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F

Notes:

**Serial #:** \_\_\_\_\_ **Optics Check:**  Pass  Fail

		Level 1					Level 2					Repeat: <input type="checkbox"/> Level 1 <input type="checkbox"/> Level 2				
Date	Initial	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail
						<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F

Notes:

**Serial #:** \_\_\_\_\_ **Optics Check:**  Pass  Fail

		Level 1					Level 2					Repeat: <input type="checkbox"/> Level 1 <input type="checkbox"/> Level 2				
Date	Initial	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail
						<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F

Notes:

**Serial #:** \_\_\_\_\_ **Optics Check:**  Pass  Fail

		Level 1					Level 2					Repeat: <input type="checkbox"/> Level 1 <input type="checkbox"/> Level 2				
Date	Initial	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail	CHOL	HDL	TRG	GLU	Pass/Fail
						<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F					<input type="checkbox"/> P / <input type="checkbox"/> F

Notes:

### Problem and Corrective Action Log:

For any questions or concerns, call your event manager, follow the user guides and package inserts provided with the products or call PTS Diagnostics Customer Service: 1 (877) 870-5610 or +1 (317) 870-5610 for toll-free calls

